

3.0 510(k) SummaryPage 1 of 1

Sponsor: Synthes (USA)
1301 Goshen Parkway
West Chester, PA 19380
(610) 719-6940

Device Name: Synthes LCP Distal Femur Plates

Classification: Class II, §888.3030 – Single/multiple component metallic bone fixation appliances and accessories.

Predicate Device: Synthes LCP Curved Condylar Plates

Device Description: Synthes LCP Distal Femur Plates are part of the Synthes Locking Compression Plate (LCP) System. The plates have a low profile design, are available in left and right versions with Dynamic Compression Plate (DCP) holes combined with locking screw holes in the shaft of the plate and threaded screw holes in the head of the plate. The plates will be offered in stainless steel and titanium and will be available in sterile and non-sterile versions.

Intended Use: Synthes LCP Distal Femur Plates are intended for buttressing multifragmentary distal femur fractures including: supracondylar, intra-articular and extra-articular condylar, periprosthetic fractures and fractures in normal or osteopenic bone, nonunions and malunions, and osteotomies of the femur.

Substantial Equivalence: Information presented supports substantial equivalence.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Synthes (USA)
% Ms. Sheri L. Musgnung
Sr. Regulatory Affairs Specialist
1301 Goshen Parkway
West Chester, Pennsylvania 19380

OCT 19 2006

Re: K062564

Trade/Device Name: Synthes LCP Distal Femur Plates
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: Class II
Product Code: HRS
Dated: August 30, 2006
Received: August 31, 2006

Dear Ms. Musgnung:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set

Page 2 – Ms. Sheri L. Musgnung

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0210. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson

Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure



2.0

Indications for Use

510(k) Number (if known): _____

Device Name: _____

Synthes LCP Distal Femur Plates

Indications for Use:

Synthes LCP Distal Femur Plates are intended for buttressing multifragmentary distal femur fractures including: supracondylar, intra-articular and extra-articular condylar, periprosthetic fractures and fractures in normal or osteopenic bone, nonunions and malunions, and osteotomies of the femur.

Prescription Use X
(Per 21 CFR 801.109)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

510(k) Number _____

1262564

000004